

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100392

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/25/2011

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 28.03: 0.074  
LOT #: 02511080A1  
EXPIRATION: 02/01/2013

BLANK TEST	0.000	14:21
INTERNAL STANDARD	VERIFIED	14:21
EXTERNAL STANDARD	0.075	14:21
BLANK TEST	0.000	14:22
EXTERNAL STANDARD	0.074	14:22
BLANK TEST	0.000	14:23
EXTERNAL STANDARD	0.075	14:23
BLANK TEST	0.000	14:24
EXTERNAL STANDARD	0.074	14:24
BLANK TEST	0.000	14:25
EXTERNAL STANDARD	0.075	14:25
BLANK TEST	0.000	14:26

Average = 0.0746  
Std Dev = 0.0005

## Diagnostic Check

### VERSIONS

DMT: 1.01  
PIC: 3.02  
Modem: 2.0  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.2°C PASSED  
Breath Tube = 46.3°C PASSED

### PUMP INFO

Flow Rate = 5.109 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED


### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 4 day of Nov, 20 11

  
Nikki Roth, Notary Public  
My Commission Expires With Office



BMB  
10/25/11